

(d) if so, what are the details thereof?

THE MINISTER OF HEALTH AND FAMILY WELFARE (SHRI B. SHAN-KARANAND): (a) No, Sir.

(b) to (d) Do not arise.

Non-functional equipments in the hospitals of Delhi

2349. PROF. VIJAY KUMAR MALHOTRA: Will the Minister of HEALTH AND FAMILY WELFARE be pleased to state:

(a) number of equipments lying non-functional in Government hospitals in Delhi equipment-wise and hospital-wise,

(b) since when these equipments are lying non-functional and the reasons for not making them functional; and

(c) what is the effect of these equipments on the efficient health care of the patients?

THE MINISTER OF HEALTH AND FAMILY WELFARE (SHRI B. SHAN-KARANAND): (a) to (c) A statement in respect of Central Government hospitals in Delhi is enclosed.

Statement

Name of Major equipment lying non-functional	Since when the equipments lying non-functional and reasons for not making them functional	Effects on the patients due to non-functioning of equipments.
1	2	3
<i>Safdarjung Hospital:</i>		
1. Weck Operating Microscope	For last one year. Non response of the suppliers to several calls for repair.	Patient care is not affected. Work being carried out with the help of other equipments.
2. Haig Street Slit lamp	Do.	Do.
3. MMR Cannon Camera	From last one year. Unit is under the process of repair	Do.
4. Cell Dyn 900 Haematology Analyser	From last one year. Under the process of repair.	Do.
5. EMG Machine	For more than a year. Under the process of repair.	Do.
6. Blood Gas analyser ABL-4	Spare parts are not available in India. Efforts are being made to procure spare parts from abroad for its repair.	Do.
7. 7200 Microprocessor	Could not be commissioned as the marketing firm kept on changing. Efforts are on, and the negotiation to install it is going on with the present agent M/s. Indian Oxygen Ltd.	Do.

1

2

3

Dr. Ram Manohar Lohia Hospital:

- | | | |
|---------------------------------------|---|---|
| 1. C. Arm Image Intensifier | Non functional since September, 1991. Parts of this imported equipment are not available. local dealers has been approached for parts. | An alternative Image Intensifier of the X-ray department is used |
| 2. Ultrasound Machine | Non functional since December, 1993. The procurement action for replacement of one transducer is in progress. | The work is being done on the other machine. |
| 3. DSA System | Non functional since September, 1992. Parts of this equipment are not available in India. Correspondence has been made in this regard to procure parts. | The engiography in DSA system are now performed on the conventional system. |
| 4. Argon Laser Photo Coagulation—900; | Non functional since November, 1990. The procurement process of Argon laser Tube for getting it imported is in progress. | Patients requiring this services are referred to other Government hospitals having this facility. |
| 5. CUSA Machine | Non-functional since, 1991. The procurement process for imported parts of this machine is in process. | Out of three functions that the machine performs only one function namely 'Vibration' is not working. |

Lady Harding Medical College & Associated Hospitals:

- | | | |
|-----------------------|--|------------------------------|
| 1. Toches—cardiograph | Non functional since 1991. Steps have been taken to get it repaired. | Patient care is not affected |
|-----------------------|--|------------------------------|

NOTE: Machine which are very old and unserviceable are in the process of being condemned.

Dental clinics in certain colonies of Delhi

2350. PROF. VIJAY KUMAR MALHOTRA: Will the Minister of HEALTH AND FAMILY WELFARE be pleased to state:

(a) whether there is any proposal under Government's consideration to establish dental units @ one each in Government colonies in Delhi like R. K. Puram, Moti Bagh to provide dental care at

the door steps of the employees;

(b) if so, the details thereof; and

(c) if not, the reasons therefor?

THE MINISTER OF HEALTH AND FAMILY WELFARE (SHRI B. SHANKARANAND): (a) No, Sir.

(b) Does not arise.

(c) Dental treatment facility is available in several hospitals located in different parts of the city.